

MECHANICAL

TEMPERATURE

PORT LOCATION: 12S SOLDER TYPE: SAC305

OPERATING: SENSITIVITY AT 500 Hz WILL NOT

-17°C TO 63°C.


STORAGE: -40°C TO 63°C.

VARY MORE THAN +1 / -3 dB FROM

AN UNDAMPED MAGNETIC BALANCED ARMATURE RECEIVER WITH AN EXTENDED HIGH FREQUENCY RESPONSE INTENDED FOR USE IN CIC HEARING INSTRUMENTS.

NOTE: SPECIFICATIONS FOLLOWED BY AN ASTERISK (*) ARE 100% TESTED.





FREQUENCY IN HERTZ

ACOUSTICAL

SENSITIVITY*

FOR

20 µPa

0

DEVICE WILL PRODUCE THE SPL LISTED BELOW UNDER TEST CONDITION IN TABLE 3. NOMINAL SENSITIVITY AT IKHZ IS dB RELATIVE TO 20 μPα. ALL OTHER VALUES IN dB RELATIVE TO THE SENSITIVITY AT IKHz.

| FREQUENCY (Hz) | MINIMUM | NOMINAL | MAXIMUM |
|----------------|---------|---------|---------|
| 200 | - 3 . 0 | 0.0 | + 3 . 0 |
| 500 | -2.0 | 0.0 | +2.0 |
| 1000 | -2.0 | 95.0 | +2.0 |
| 3000-3600 | +5.5 | +8.5 | +11.5 |
| 5800-6800 | -1.5 | +1.5 | +4.5 |

TABLE I

TOTAL HARMONIC DISTORTION*

DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

| FREQUENCY (Hz) | AC DRIVE (V rms) | DC BIAS (mA) | LIMIT (%) |
|----------------|------------------|--------------|-----------|
| 1058 | 0.211 | 0 | 5 |
| 1588 | 0.211 | 0 | 5 |
| 1588 | 0.595 | 0 | 10 |

TABLE 2

TEST CONDITIONS

| NOMINAL SOURCE VOLTAGE | 0.211 V rms, 0 mA DC BIAS |
|------------------------|---|
| SOURCE IMPEDANCE | < I Ohm |
| TUBING | |
| COUPLER CAVITY | 2 CM ³ , SIMULATED ANSI S3.7 TYPE HA-3 (IEC 60318-5) |

TABLE 3

ELECTRICAL

| DC RESISTANCE | 100 Ohms ± 10%* |
|--------------------|-----------------|
| IMPEDANCE @ 500 Hz | Ohms ± 5%* |
| IMPEDANCE @ IkHz | 124 Ohms ± 15%* |

TABLE 4

ISOLATION: CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT.*

| Revision | C.O. # | Implementation Date | KELEASE LEVEL | | REVISION. |
|--|-----------|---------------------|-------------------|--------------------|-----------|
| D | C10114032 | -8- 2 | Active | | D |
| WHEN TEST LIMITS ARE USED TO ESTABLISH INCOMING INSPECTION ACCEPTANCE/REJECTION CRITERIA, CORRELATION OF TEST EQUIPMENT WITH KNOWLES IS ALSO REQUIRED FOR ELIMINATION OF EQUIPMENT AND TEST METHOD VARIATION | | | DR. BY MMM CK. BY | DATE 6 - 3 - 0 6 | |
| TITLE: | RE | CEIVER | WBFK-30000-000 | GJP | 6-23-06 |

SHT 2.1

6-23-06

PERFORMANCE SPECIFICATION

KNOWLES ELECTRONICS ITASCA, ILLINOIS U.S.A.

Mouser Electronics

Authorized Distributor

Click to View Pricing, Inventory, Delivery & Lifecycle Information:

Knowles:

WBFK-30000-000