

We measure it.



Differential pressure measuring instrument

testo 510 – Pocket-sized differential pressure measurement

Differential pressure measurement 0 to 100 hPa

Flow velocity measurement possible with Pitot tube

Temperature and air density compensation

Display illumination

10 selectable units



hPa

Illustration 1.1

testo 510 measures differential pressure in the range from 0 to 100 hPa. The differential pressure measurement is temperature-compensated for accurate measurement values. The measurement values can be displayed in Pascal over the entire measurement range. Magnets at the rear permit free-hand work. The backlit display allows the

measurement values to be easily read out, even in unfavourable light conditions. The testo 510, in combination with a Pitot tube, measures air flow velocity. For accurate measurement values, the air density can be compensated. testo 510 is very handy, small and easy to operate.

Technical data / Accessories

testo 510

testo 510 handy measuring instrument for differential pressure incl. protection cap, batteries, belt holder and calibration protocol



Part no. 0560 0510

Sensor type	Differential pressure sensor
Measuring range	0 to 100 hPa
Accuracy ±1 digit	±0.03 hPa (0 to 0.30 hPa) ±0.05 hPa (0.31 to 1.00 hPa) ±(0.1 hPa + 1.5 % of mv) (1.01 to 100 hPa)
Resolution	0.01 hPa

General technical data	
Positive pressure	500 mbar
max. static pressure	1.5 bar
Operating temperature	0 to +50 °C
Storage temperature	-40 to +70 °C
Selectable units	hPa, mbar, Pa, mmH2O, inH2O, inHg, mmHg, psi, m/s, fpm
Protection class	IP40
Battery type	2 AAA micro batteries
Battery life	50 h (average, without display illumination)
Measurement rate	0.5 s
Dimensions	119 x 46 x 25 mm
Weight	90 g (with batteries and protective cap)
Warranty	2 years

Accessories	Part no.
Accessories for measuring instrument	
Connection hose, silicone, 2 m long, max. load 700 hPa (mbar)	0554 0448
Belt holder	0516 4007
ISO calibration certificate pressure; differential pressure; 3 points distributed over meas. range	0520 0095
ISO calibration certificate pressure; differential pressure; 5 points distributed over meas. range	0520 0005

0981 9764/msp/A/10.2013

Subject to change without notice.

