



Quality and reliability is our tradition

KYORITSU

AC DIGITAL CLAMP METER KEW 2200

The Slim and Lightweight 1000A AC Clamp Meter



- **Ultra Slim and lightweight Handy design**
- **1000A AC Clamp Meter**
AC40/400/1000A Auto Range
- **Ø33mm Tear Drop Jaw easy to use in tight places**
- **DMM function ACV, DCV, Ω , Continuity Buzzer**
- **Fuseless electronic protection on Ω up to 600V**
- **IEC61010-1 Safety Standard**
CAT.Ⅲ 600V (AC A)
CAT.Ⅲ 300V/CAT.Ⅱ 600V (AC/DC V)

Actual Size



KYORITSU ELECTRICAL INSTRUMENTS WORKS,LTD.

<http://www.kew-ltd.co.jp>

● KEW 2200 Specifications

AC A	40.00/400.0/1000A (Auto-ranging) ±1.4%rdg±6dgt(50/60Hz) ±1.6%rdg±6dgt(45~65Hz)
AC V	4.000/40.00/400.0/600V (Auto-ranging) ±1.8%rdg±7dgt(45~65Hz) ±2.3%rdg±8dgt(65~500Hz)
DC V	400.0mV/4.000/40.00/400.0/600V (Auto-ranging) ±1.0%rdg±3dgt
Ω	400.0Ω/4.000/40.00/400.0kΩ/4.000/40.00MΩ (Auto-ranging) ±2.0%rdg±4dgt(0~400kΩ) ±4.0%rdg±4dgt(4MΩ) ±8.0%rdg±4dgt(40MΩ)
Continuity buzzer	buzzer sounds below 50±30Ω
Conductor size	φ33mm max.
Withstand voltage	5320V AC for 5 seconds(AC A) 3540V AC for 5 seconds(AC/DC V)
Applicable standards	IEC 61010-1 CAT.Ⅲ 600V (AC A), CAT.Ⅲ 300V/CAT.Ⅱ 600V(AC/DC V), Pollution degree 2 IEC 61010-031, IEC 61010-2-032 IEC 61326(EMC)
Power source	R03/LR03(AAA)(1.5V) × 2
Dimensions	190(L) × 68(W) × 20(D)mm
Weight	Approx. 120g(including batteries)
Accessories	7107A(Test leads), 9160(Carrying case), R03(AAA)× 2, Instruction manual
Optional	8008(Multi-tran)

● Included Accessories



● Optional Accessories

MODEL 8008
Multi-tran



For measurements
up to 3000A AC



Teardrop shaped jaw for easy use in crowded cable areas



Capable of measuring currents up to 1000A AC




Slim and compact that fits right into your pocket



Standard 4 mm input terminals for ordinary test leads


AC Current Measurement

Press the trigger to open the current sensor and clamp the one conductor under test.



AC/DC Voltage Measurement

Connect the test leads to the circuit under test.



Safety Warnings :

Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

■ For inquiries or orders :



**KYORITSU ELECTRICAL
INSTRUMENTS
WORKS, LTD.**

No.5-20,Nakane 2-chome, Meguro-ku, Tokyo, 152-0031 Japan

Phone:81-3-3723-0131

Fax:81-3-3723-0152

E-mail:info@kew-ltd.co.jp

Factory: Ehime, Japan

<http://www.kew-ltd.co.jp>



ISO9001:2008
APPROVED BY BUREAU VERITAS Certification



In consideration of the environment, soy ink and recycled paper were used in this publication. ● The contents of this leaflet are subject to change without notice.

KEW 2200-1E Mar. 11 AD